

State of California—Health and Human Services Agency California Department of Public Health



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Philips Respironics recalls V60 Series Ventilators for Internal Electrical Circuit Issue

Recall Date	Product Description	Recalling Firm	Recall Reason
2/28/2022	Philips Respironics V60 Ventilator (All Models)	Respironics California, LLC Carlsbad, California	Ventilator units have been identified to have an issue impacting an internal electrical circuit within the ventilators, that in some case could potentially result in the ventilator ceasing to operate.
2/28/2022	Philips Respironics V60 Plus Ventilator (All Models)	Respironics California, LLC Carlsbad, California	Internal Electrical Circuit Issue

Recall Class	Product Identification	Distribution	Affected Dates
I	Philips Respironics V60 Ventilator All Lots & Serial Numbers	Nationwide including California	February 2022 and prior
I	Philips Respironics V60 Plus Ventilator All Lots & Serial Numbers	Nationwide including California	February 2022 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE FDA WEBSITE

CDPH Food and Drug Branch MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435 (916) 650-6500 ● (916) 650-6650 FAX Internet Address: <u>www.cdph.ca.gov</u>

