



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

3M Recalls Red Dot Electrode due to Breakage Potential

Recall Date	Product Description	Recalling Firm	Recall Reason
11/10/2021	3M Red Dot Electrodes, Radiolucent Monitoring Electrode with Foam Catalog No. 2244 3M ID (SKU): 70-2011-8506-6 SAP ID: 7100212989 packaged 10 electrodes per card, 5 cards per pouch, 20 pouches per shipper box	3M Company - Health Care Business St. Paul, Minnesota	Potential for breakage of the ECG electrode carbon stud from the eyelet.

Recall Class	Product Identification	Distribution	Affected Dates
II	3M Red Dot Electrodes Lot #202302LK, Exp. 2023-02-19 pouch UDI (01)1070738789449(17)230219(11)202302LK shipper box UDI (01)5070738778 9447(17)230219(11)202302LK Lot #202302LL, Exp. 2023-02-22 pouch UDI (01)1070738789449(17)230222(11)202302LL shipper box UDI (01)5070738778 9447(17)230222(11)202302LL Lot #202302LM, Exp. 2023-02-22, pouch UDI (01)1070738789449(17)230222(11)202302LM shipper box UDI (01)5070738778 9447(17)230222(11)202302LM Lot #202302LN, Exp. 2023-02-23. pouch UDI (01)1070738789449(17)230223(11)202302LN shipper box UDI (01)5070738778 9447(17)230223(11)202302LN.	40 cases in California	October 2021 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

