



CALIFORNIA DEVICE RECALL INFORMATION SHEET

Karl Storz Endoscopy Recalls Videoscope due to Infection Risk

Recall Date	Product Description	Recalling Firm	Recall Reason
11/10/2021	FIVE S 3.5x65, Flexible Intubation Videoscope REF: 091361-06 (6 Units)	Karl Storz Endoscopy USA El Segundo, California	Affected flexible intubation video endoscopes did not achieve the required sterility assurance level. There is a risk that the patient may be exposed to a higher risk of infection.

Recall Class	Product Identification	Distribution	Affected Dates
II	FIVE S 3.5x65, Flexible Intubation Videoscope All LOTs with remaining shelf life	330 Intubation Scopes in California	September 2021 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

