



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

TiSport Recalls Aero Z and ZRA Wheelchairs with Faulty Tiedowns

Recall Date	Product Description	Recalling Firm	Recall Reason
09/24/2021	Permobil TiLite Models: Aero Z, and ZRA	TiSport LLC, subsidiary of Permobil Pasco, Washington	Some affected wheelchairs, with transit tie downs, that are used as a passenger seat in a motor vehicle may not be configured in the right orientation to perform as intended if involved in a motor vehicle accident.

Recall Class	Product Identification	Distribution	Affected Dates
II	M2005054 to 1231112011	461 Units in California	January 2017 – June 2021

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
 (916) 650-6500 • (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

