



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Smith and Nephew Recalls EVOS Self-Tapping Locking Screw SST**

Recall Date	Product Description	Recalling Firm	Recall Reason
10/25/2021	<b>EVOS Self-Tapping Locking Screw SST</b>  REF 72413575, 3.5MM x 75MM REF 72413570, 3.5MM X 70MM. Sterile.	<b>Smith &amp; Nephew, Inc.</b> Memphis, Tennessee	The labels on the packages of screws were switched.

Recall Class	Product Identification	Distribution	Affected Dates
II	REF 72413575 - batch #21FM24289 REF 72413570 - batch #21GM04743.	Nationwide including California	October, 2021 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch  
 MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435  
 (916) 650-6500 ● (916) 650-6650 FAX  
 Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

