

State of California—Health and Human Services Agency

California Department of Public Health



Food and Drug Branch - Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Whele Recalls Mighty Bliss Electric Heating Pad for Potential Burns

Recall Date	Product Description	Recalling Firm	Recall Reason
11/30/2022	Mighty Bliss Electric Heating Pad Large (12 x 24) Blue Electric Heating Pad, MB-001	Whele LLC DBA Perch Boston, Massachusetts	The products may have the potential to burn customers. Firm has received complaints of burns, mild shocks, and rashes/irritation.
11/30/2022	Mighty Bliss Electric Heating Pad Extra-Large (20 x 24) Blue Electric Heating Pad, MB-002	Whele LLC DBA Perch Boston, Massachusetts	The products may have the potential to burn customers. Firm has received complaints of burns, mild shocks, and rashes/irritation.
11/30/2022	Mighty Bliss Electric Heating Pad Large (12 x 24) Grey Electric Heating Pad, PE-MtyBls-HeatPad- 12x24-Gry-V2	Whele LLC DBA Perch Boston, Massachusetts	The products may have the potential to burn customers. Firm has received complaints of burns, mild shocks, and rashes/irritation.

Recall Class	Product Identification	Distribution	Affected Dates
I	12x24 Blue No UDI Lots 210108- 220103	528,067 Nationwide	October 2022 & prior
I	20x24 Blue No UDI Lots 210901 211001	12,140 Nationwide	October 2022 & prior
I	12x24 Grey No UDI Lot 220103	4,034 Nationwide	October 2022 & prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE FDA WEBSITE

